

Table 2-4: Analytical Parameters

Parameter	EPA Method	Technique	Detection Limit (ug/L)
Copper	200.8	ICPMS	0.5
Lead	200.8	ICPMS	0.5
Nickel	200.8	ICPMS	0.5
Zinc	200.8	ICPMS	0.5
TDS	160.1	Gravimetric	1
TSS	160.2	Gravimetric	1

Table 2-5: Field Water Quality Measures

Parameter	Technique
Depth	Measuring cable
pH	Horiba U10
Conductivity	Horiba U10
Dissolved Oxygen	Horiba U10
Turbidity	Horiba U10
Temperature	Horiba U10
Salinity	Horiba U10