
State Water Resources Control Board

April 13, 2021

Dr. Agustin Pierri
Technical Director
Weck Laboratories, Inc.

Dear Dr. Pierri,

This letter serves as notice that Weck Laboratories, Inc. has been approved for the use of the following methods for the purposes of monitoring for constituents of emerging concern pursuant to the Water Quality Control Policy for Recycled Water (Recycled Water Policy):

- Method EPA 8270 for the analysis of 1,4-Dioxane
- Method EPA 521 for the analysis of NDMA and NMOR
- Method EPA 537.1 for the analysis of PFOS and PFOA
- Method EPA 1694M for the analysis of Gemfibrozil, Iohexol, Sucralose, and Sulfamethoxazole

West Basin Municipal Water District submitted a validation package on behalf of Weck Laboratories, Inc. for these methods on March 31, 2021. This validation package demonstrated the utility of the above methods for the respective analytes in a water quality representative of the matrices outlined in Attachment A of the Recycled Water Policy (i.e., West Basin Barrier Product water).

This approval for the use of the above methods for the respective analytes is only for the purposes outlined in the Recycled Water Policy, Attachment A, and this approval does not extend to the use of these methods for any other purpose.

Sincerely,



Tricia Lee

cc:

Laura McLellan, State Water Resources Control Board
Steven Webb, Los Angeles Regional Water Quality Control Board
Jamie Burkhard, West Basin Municipal Water District
Uzi Daniel, West Basin Municipal Water District
Jamie Malpede, West Basin Municipal Water District